



HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED
Procurement Cell
(A Government of Maharashtra Undertaking)
Regd. Office : Acharya Donde Marg , Parel, Mumbai 400 012 (INDIA)

| | |
|--|--|
| Phone No: 022- 24129320-23 Managing Director :022-24150628 General Manager (Procurement Cell) :022-24100478 | Website : http://www.vaccinehaffkine.com E-mail: procurementcell@vaccinehaffkine.com |
| प्रशासकीय मंजूर निधी १७,३८,०६९/- (सन २०१९-२०) पंचवार्षिक योजना, राज्य योजना | No.: 6155 /Haffkine/Procurement Cell/E-2341/ Cautery (Surgical Diathermy)/2021-22. Date: 12-01-2022 |

To,
M/s. Alan Electronic Systems Pvt. Ltd.
F-75, MIDC, Additional Ambernath,
Indl. Area, Anand Nagar, Jambivli,
Ambernath (East)
Dist: Thane-421506.
Maharashtra, India.
E-Mail: feedback@alanindia.net

Subject : Supply Order for Tender No. E-2341/Cautery (Surgical Diathermy).

- Reference:**
1. Tender No. E-2341/HBPCL/PC/Cautery (Surgical Diathermy)/2019-20.
 2. शा. नि. क्र. : खरेदी-२०१९/प्र.क्र.१०८/आरोग्य-८ दि. १८ सप्टेंबर, २०१९.
 3. **Sanction of Tender Approval Committee Meeting Dated 22.09.2021.**

With reference to the tender cited under reference no 1, you are requested to supply the following goods as per details mentioned below to consignee list enclosed with this order.

| Sr. No. | Name of the item | Specification of item | Quantity/ Unit (DHS) | Unit Rate inclusive of GST(Rs.) | Total Amount Rs. |
|---|--|--------------------------|----------------------|---------------------------------|------------------|
| 1. | Cautery (Surgical Diathermy) Make : "ALAN" Model : ELSY-360L+ | As per Annexure X | 05 | 98,704/- | 4,93,520/- |
| Total amount in words: Rupees Four Lakhs Ninety Three Thousand Five Hundred Twenty Only. | | | | | |
| Factory Location: M/s. Alan Electronic Systems Pvt. Ltd. F-75, MIDC, Additional Ambernath, Indl. Area, Anand Nagar, Jambivli, Ambernath (East) Dist: Thane-421506. Maharashtra, India. | | | | | |

- 1 **Forwarding:** Forwarding Free on Road Destination. I.e. door delivery basis.
- 2 **Delivery Period:** 06 weeks from the date of receipt of order by the supplier to the consignee attached.
- 3 **Pre-Dispatch Inspection:** Supplier shall make necessary arrangement / facilitate to carry out Pre-Dispatch inspection as per Tender Terms & condition and submit the Inspection report to this office. The Pre-Dispatch inspection cost will be borne by supplier. Machine should be dispatched only after satisfactory Pre-Dispatch Inspection.

- 4 **Risk purchase clause:** If the bidder fails to supply the stores within the stipulated delivery period, the order will stand cancelled. Undersigned shall be entitled to purchase such stores from any other source at such price which ordinarily should not be more than 10% of the tender price. The extra expenditure in such cases shall be recovered by Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai from the Supplier.
- 5 **Payment Terms:** Payment of 100% of the contract value will be made within 8 weeks on delivery and successful installation and satisfactory commissioning and operation of the machinery.
- 6 **Acceptance & Receipt:** It should be submitted in Appropriate Format to the purchasing authority.
- 7 **Delivery Challan** - Should be sent in the name of consignee in duplicate. It should specify Name of Equipment / Mfg. by / packing & quantity.
- 8 **Invoice Copy** - Should be sent in triplicate on the Name of Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd.(Procurement Cell), Mumbai. Along with Bill of Entry and Country of Origin Certificate of the consignment.
- 9 **Other Terms :**
- 1) Warranty: The warranty period shall be for 2 years from the date of commissioning of all equipment supplied as certified by the consignee. After completion of 2 years warranty period Manufacturer/Supplier should give commitment to ensure services and supply of spare part for further 8 years. The successful tenderer must ensure 95% uptime during warranty period. In case of downtime, warranty period will be extended for period of downtime. If the equipment is not attended within 24 hours for Mumbai and 48 hours for other places the supplier will be liable to pay a penalty of 0.07% of purchase cost for every day of delay. Such penalty will be recovered from the amount of security deposit. Certificate of such uptime / downtime issued by the end user will be binding for the supplier Replacement of spares parts thereof due to manufacturing defects during warranty period will be entirely at the supplier's cost.
- 2) The user institution will enter to the Comprehensive Maintenance Contract with supplier agency @ 5% of the order value (excluding taxes) of the equipment per year for 8 years after completion of warranty period. In case of non-compliance of CMC the supplier will be liable to pay penalty or for appropriate action. Payment of CMC on yearly basis will be made by the user's institution, at the end of the year after satisfactory performance report from the end user.
- 10 **Contract Agreement:** Bidder should submit Contract Agreement on non-judicial stamp paper of requisite value.

Fall Clause

It is a condition of the contract that all through the currency thereof, the price at which you will the supply stores should not exceed the lowest price charged by you to any customer during the currency of the contract and that in the event of the prices going down below the rate contract prices you shall promptly furnish such information to us to enable to amend the contract rates for subsequent supplies.

- 11 The Bidder should submit (within 7 days) amount of 1.5% i.e. **Rs. 74,028/-** of order value to meet other incidental expenditure and 3% i.e. **Rs. 1,48,056/-** as Security Deposit in form of Bank Guarantee. The Bank Guarantee valid for 2 months after the expiry of date of warranty issued by any Nationalized / Scheduled Bank.

Amount to be deposited to Following Account:

| | |
|---------------------------|--|
| Name of Account | Haffkine Bio-Pharmaceutical Corporation Ltd. (Procurement Cell), CESS Account Mumbai. |
| Name of the Bank & Branch | Bank of Maharashtra, Branch- Mumbai Parel |
| Account No. | 60381379835 |
| IFSC Code | MAHB0000079 |

Consignee: As per list enclosed

ग. व्यवस्थापकीय संचालक यांच्या मान्यतेने व करिना


Smt. Sushama Patil
General Manager

Haffkine Bio Pharmaceutical Corporation Ltd.
(Procurement Cell), Mumbai.

Copy to:

- 1) Commissioner Health Services, Mumbai.
- 2) Director, Medical Education & Research, Mumbai-400 001.
- 3) Account Manager, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.
- 4) Office File.

Copy to Consignee: As per Consignee List.: As per Tender Condition No.17 The user Institution should get the Comprehensive Maintenance Contract done with supplier agency @ 5% of the Order value (excluding taxes) of equipment per year for Eight years after Completion of warranty period.

Copy Submitted to: 1) Secretary, Medical Education & Drug Department, Mantralaya, Mumbai.

Annexure-X

| Sr. No. | | Technical Specifications |
|----------------|--|---|
| 1 | Clinical Purpose | Cautery uses an electric current to produce heat deep inside a targeted tissue. It can reach areas as deep as two inches from the skin's surface. The diathermy machine does not apply heat directly to the body. Instead, the current from the machine allows the body to generate heat from within the targeted tissues |
| 2 | Used by clinical department/ward | Operation theater |
| 3 | Technical characteristic specific to this type of device | <ol style="list-style-type: none"> 1. Facility for Monopolar , Bipolar and underwater cutting. Monopolar cutting & coagulation 2. Micro-processor based technology 3. Monopolar cut in minimum 3 modes 4. Bipolar-coagulation in 3 or more modes (forced coagulation, spray coagulation and soft coagulation). 5. Blending of cutting and coagulation-in minimum 2 levels. 6. Automatic cut-off technology with self check on every start. 7. Foot and hand switch. 8. Auto monitoring and display of set parameters. 9. Digital or Touch-controlled interface to set parameter. 10. Four or more programmable memory. 11. Simultaneous use of Monopolar and Bipolar Coagulation. 12. Output Power of 400 Watt (Minimum) 13. Monopolar Cutting- 400Watt and Coagulation power adjustable from 0-150 Watt or more. 14. Bipolar Coagulation power adjustable from 0-50 W, Micro Power Range-0.1 to 9.9 Watt increment of 0.1 Watt, Macro power range from 1-50Watt increment of 1 Watt. 15. Audio-visual Alarm for disconnection of Neutral Plate. |
| 4 | User's interface | Manual |
| 5 | Software and/or Standard of communication | In-Built |
| 6 | Weight (lbs, kg) | Max: 10 kg |
| 7 | Heat dissipation | Heat Dissipation: Should Maintain nominal temp and the heat should be disbursed through an cooling mechanism. |
| 8 | Mobility, Portability | Portable |
| 9 | Power Requirements | Recharging unit Input voltage – 220V-240V Ac, 50Hz |
| 10 | Tolerance (to variations, shutdowns) | ±10% |
| 11 | Protection | Should have over-charging cut-off with visual symbol |
| 12 | Power consumption | 60W |
| 13 | Accessories (mandatory) | <ol style="list-style-type: none"> 1. Power cord: 1pc 2. Electrode lever: 1pc 3. Electrode: 2sets 4. Reusable electrode handle with cutting/coagulation switch 5. Disposable REM plate. 6. Cable for electrode handle. 7. Neutral plate for adults and pediatric |
| 14 | Atmosphere/ Ambiance (air conditioning, | <ol style="list-style-type: none"> 1. Operating condition : Capable of operating continuously in ambient temperature of 10 to 40°C and relative humidity of 15 to 90% in ideal circumstances. |

| | | |
|----|--|--|
| | humidity, dust) | 2. Storage condition : Capable of being stores continuously in ambient temperature of 0 to 50°C and relative humidity of 15 to 90 %. |
| 15 | User's care, Cleaning, Disinfection & Sterility issues. | 1. Disinfection : Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/ disposable cover. 2. Sterilization not required. |
| 16 | Certificates | 1. Should be FDA/CE/BIS approved product. 2. Manufacturer and Supplier should have ISO certification for quality standards. 3. Electrical safety conforms to the standards for electrical safety. IEC 60601- General requirements (Or equivalent BIS Standard) 4. Certificates with test reports for electrical safety. |
| 17 | Local and/or international | Manufacturer /supplier should have ISO 13485 certificate for quality standard. |
| 18 | Pre-installation requirements: nature, values, quality, tolerance | 1. Availability of 5 amp socket. 2. Safety and operation check before handover. |
| 19 | Requirements for sign-off | Certificate of calibration and inspection from the manufacturer. |
| 20 | Training of staff (medical, paramedical, technicians) | 1. Training of users on operation and basic maintenance. 2. Advanced maintenance tasks required shall be documented. |
| 21 | Warranty | 1. Two year on site with free servicing (min.03/year) during warranty & provide technical support and required spares and consumable for 7 yrs after warranty period. 2. In warranty period for the repair if required more than 24 hours, provide stand by same cautery machine. |
| 22 | Maintenance tasks | 1. Maintenance manual detailing. 2. Complete maintenance schedule. |
| 23 | Service contract clauses, including prices | The same price list of all spares and accessories (including minor) required for maintenance and repairs in future after guarantee/warranty period should be attached. |
| 24 | Operating manuals, service manuals, other manuals | Should provide 2 sets (hardcopy and soft copy) 1. Users, technical and maintenance manuals to be supplied in English/Marathi Hindi language along with machine diagrams 2. List of equipment and procedures required for local calibration and routine maintenance. 3. Service and operation manuals (original and copy) to be provided. 4. Advanced maintenance tasks documentation. 5. Certificate of calibration and inspection. |
| 25 | Other accompanying documents | List of important spares and accessories, with their part numbers and cost. |
| 26 | Service Support Contact details (Hierarchy wise ; including a toll free/landline number) | Contact details of manufacturer, supplier and local service agent to be provided, Any Contract (AMC/CMC/add-hoc) to be declared by the manufacturer |
| 27 | Recommendations or warning | Any warning signs would be adequately display. |

Consignee Details

| | | |
|--|--|-------------|
| M/s. Alan Electronic Systems Pvt. Ltd. | | |
| Cautery (Surgical Diathermy) Make : "ALAN" Model : ELSY-360L+ | | |
| Delivery Period | 06 weeks | |
| PO Ref. No. | No.: 6155 /Haffkine/Procurement Cell/E-2341/Cautery (Surgical Diathermy) /2021-22. Date: 12-01-2022 | |
| प्रशासकीय मंजूर निधी १७,३८,०६९/- (सन २०१९-२०) पंचवार्षिक योजना, राज्य योजना | | |
| Sr. No. | Name & Address of the Consignee | Qty. |
| 1 | Rural Hospital, Jejuri, Pune. | 1 |
| 2 | Rural Hospital, Bawada, Pune. | 1 |
| 3 | Rural Hospital, Bhigwan, Pune. | 1 |
| 4 | Trauma Care Unit, Rural Hospital, Rajgurunagar, Pune. | 02 |
| | Total | 05 |

या. व्यवस्थापकीय संचालक यांच्या माध्यमेने व क्रिया


Smt. Sushama Patil

General Manager
Haffkine Bio Pharmaceutical Corporation Ltd.
(Procurement Cell), Mumbai.